

# Crystal Digital PCR® Assay

## Information Sheet

For Research Use Only. Not for use in diagnostic procedures.

### Product Name

c-KIT (D816, D816V) Crystal Digital PCR® Assay (R51010)

### Description

#### Detected Targets

Targets	Sample Type	Detection Channels	Multiplex
c-KIT (D816, D816V)	DNA	Blue/Red	2

The c-KIT (D816, D816V) Crystal Digital PCR® Assay is a 10X assay designed to detect and quantify 1 mutation in the c-KIT gene using the Ruby Chip. c-KIT encodes for a receptor tyrosine kinase that plays a crucial role in cell signaling for processes including cell survival, proliferation, and differentiation.

#### Assay configuration

The table below indicates with a “X” which channel(s) are used for each target in the assay:

Targets	Base changes	Blue	Teal	Green	Yellow	Red	Infra-Red	Long-Shift
Wild-type (WT) c-KIT D816	N/A	X						
c-KIT D816V	c2447 A>T					X		

#### Components

c-KIT (D816, D816V) Crystal Digital PCR® Assay comprises two reagents: a pool of the assay specific primers and Crystal Flex Probes and a Positive Control. Please refer to the lot specific Certificate of Conformity for characterized concentration, available for download at the Technical Resources section of the Stilla Technologies website.

Component Name	Reference	Concentration	Description
c-KIT (D816, D816V) – Crystal Digital PCR® Assay	R51010	10X	Detects 1 mutation in the c-KIT gene
c-KIT Positive Control	R51010.PC0	10X	Contains: hgDNA and synthetic c-KIT D816V mutant

## Thermocycling Programs

### On the naica® system:

Step		Ramp rate
<b>Step 1</b>	Partition for Ruby Chip	-
<b>Step 2</b>	Temperature 95°C for 180 seconds	1°C/sec
<b>Step 3</b>	Begin Loop for 60 Iterations	-
<b>Step 3.1</b>	Temperature 95°C for 15 seconds	1°C/sec
<b>Step 3.2</b>	Temperature 58°C for 30 seconds	1°C/sec
<b>Step 4</b>	Release for Ruby Chip	-

### On the Nio™ Digital PCR:

Step		Ramp rate
<b>Step 1</b>	Partition for Ruby Chip	-
<b>Step 2</b>	Temperature 95°C for 180 seconds	1°C/sec
<b>Step 3</b>	Begin Loop for 60 Iterations	-
<b>Step 3.1</b>	Temperature 95°C for 15 seconds	2°C/sec
<b>Step 3.2</b>	Temperature 60°C for 30 seconds	2°C/sec
<b>Step 4</b>	Temperature 58°C for 300 seconds	1°C/sec
<b>Step 5</b>	Release for Ruby Chip	-

## Image Acquisition

Download the dedicated scanning file from the Technical Resources section of the Stilla Technologies website:

- ScanningTemplate\_Prism3\_cKIT\_R51010.ncx (3-color naica® system)
- ScanningTemplate\_Prism6\_cKIT\_R51010.ncx (6-color naica® system)
- NioProtocol\_cKIT\_R51010.nioprotocol (Nio™ Digital PCR)
- NioAssay\_3C\_cKIT\_R51010.nioassay (Nio™ Digital PCR)

## Image Analysis

The following files are embedded in the dedicated scanning files listed above:

- CompensationMatrix\_Prism3\_cKIT\_R51010.ncm (3-color naica® system)
- CompensationMatrix\_Prism6\_cKIT\_R51010.ncm (6-color naica® system)
- CompensationMatrix\_Nio\_cKIT\_R51010.ncm (Nio™ Digital PCR)
- AnalysisConfiguration\_cKIT\_R51010.nca (all systems)

## Consumables Required but Not Provided

- Ruby Chip (C16011)
- naica® PCR MIX 10X (R10106)
- Crystal Universal Reporters 3 (R41401 200 reactions, R41402 1000 reactions)
- Nuclease-free water

## Instruction for PCR Mix Preparation


Specific instructions for preparing the PCR mix are given below.

Reagent Name	Initial Concentration	Final Concentration	Volume per reaction (µL)
naica® PCR MIX Buffer A <span style="color: green;">●</span>	10x	1x	0.60
naica® PCR MIX Buffer B <span style="color: red;">●</span>	100%	4%	0.24
Crystal Digital PCR® Assay <span style="color: orange;">●</span>	10x	1x	0.60
Crystal Universal Reporter Tube A <span style="color: yellow;">●</span>	40x	1x	0.15
Nuclease-free water	NA	NA	Variable
<b>Template DNA</b>	<b>NA</b>	<b>NA</b>	<b>Variable</b>
<i>or Positive Control</i> <span style="color: black;">○</span>	10x	1x	0.60
<i>Total reaction volume (µL)</i>			<b>6.0</b>

## Representative Data and Instructions for Analysis

Set thresholds for separating positive and negative populations on the 1D plots. To optimize the analysis, the blue and the red thresholds should be set at approximately equal distance from the positive and negative clusters. Examples of results obtained on the 6-color naica® system are given below.

Wet lab testing was carried out using genomic hgDNA and H<sub>2</sub>O as negative controls and a positive control consisting of hgDNA and synthetic c-KIT D816V mutant. Synthetic c-KIT D816V mutant was also tested individually.



**Figure 1: 1D plots obtained during wet lab testing on the 6-color naica® system.** The blue and the red thresholds should be set at approximately equal distance from the positive and negative clusters.



Stilla Technologies

F-94800 Villejuif, FRANCE

Registered names and trademarks used in this document, even when not specifically marked, are not to be considered unprotected by law.